



مستشفى الجامعة الأردنية

JORDAN UNIVERSITY HOSPITAL

دائرة العطاءات والتزويد

Tenders and Supply Department

الرقم :

التاريخ :

نموذج طلب عرض أسعار لطلب الشراء رقم (225000415)

شعبة المشتريات

السادة:-

فاكس: ()

تحية طيبة وبعد ،،،
يرجى التكرم بتقديم عرض أسعار للمواد المبينة مواصفاتها تالياً:

أسم المادة

الكمية

الوحدة

الرقم

ECG

1

eac

1

حسب المرفق

مدير دائرة العطاءات والتزويد

الصيدلانية لبنى عبد الخرابشة

على المتقدم الالتزام بالشروط التالية:-

- 1- يقدم عرض الاسعار موقع وبمغلف مختوم ومغلق أو عن طريق الايميل الرسمي Purchase.juh@ju.edu.jo او الفاكس 5300786-5526317 .
- 2- تحديد مدة التسليم وبلد المنشأ.
- 3- تقديم عينات وكتالوجات .
- 4- صلاحية العرض لا تقل عن (90) يوم .
- 5- اخر موعد لتقديم العروض يوم الاحد الموافق 01/06/2025 الساعة الثانية عشرة ظهراً.
- 6- يستبعد كل عرض معدل يدوياً.
- 7- يستبعد كل عرض غير مستوفي للشروط أعلاه
- 8- تسلم العروض في دائرة العطاءات والتزويد - شعبة المشتريات
- 9- في حال عدم توفر المواد إرسال اعتذار عن عدم توفرها على الفاكس رقم (5300786)
- 10- دفع الرسوم المقررة وتقديم تامين حسن التنفيذ خلال 10 ايام من تاريخ التبليغ ويعفى من تقديم كفالة حسن التنفيذ اذا كانت مدة التسليم لا تتعدى او تتجاوز الاسبوعين

للاطلاع على اعلانات شراء المستشفى المحلية يرجى زيارة الموقع الالكتروني www.Hospital.Ju.edu.jo

الرجاء اضافة رابط موقع الجامعة الاردنية الالكتروني (www.ju.edu.jo) على صفحاتكم الالكترونية والدخول على الموقع

تلفون: 5353666 (962-6) - فاكس : 5300786 - صندوق بريد : 13046 - عمان - الاردن

Tel : 5353666 - Fax : 5300786 - P.O.Box : 13046- Amman - Jordan

Specifications for Rest ECG recorder

1. Lightweight, easy to use and portable resting ECG recorder that can acquire a real time 12 lead interpretive ECG with user selectable 3/4 or 6 channels displayed simultaneously by manual or auto mode.
2. Built in thermal printer with a selectable chart speed of 10, 25 or 50 mm/s using A4 type paper, it should generate a report that include annotations such as date , time , patient data , medications and instrument setting.
3. Provided with alphanumeric keypad to enter patient data, date, time, etc
4. User selectable lead sensitivity of 5, 10 or 20 mm/v with a frequency range of 0.05 – 150 Hz or better.
5. Provided with original trolley and ECG patient cable hanger integrated with articulation arm.
6. Large LCD TFT type display.
7. Durable rechargeable battery with visual AC power, battery status and low battery indicators.
8. Configurable for neonate patient.
9. Visual / Audible indication of lead off / fault detection
10. 220 V / 50 Hz power supply / rechargeable battery operation (at least one hour of continuous operation when fully charged), provided with automatic battery charging function when the unit is plugged into the line power.
11. Immunity against EMG artifacts and power line noise (50 / 60 Hz noise) using digital filters.
12. Provided with the required accessory kit to operate the unit such as:
 - ECG patient cable CLIP, qty (2).
 - Power cord, leg-stand type, qty (1).
 - Recording paper, qty (12).
 - Any other recommended spare parts and accessories can be appended with the company's offer.

13. It should hold FDA clearance / CE mark.

14. At least two years comprehensive warranty should be assigned the proposed unit.

15. Operation manual and advanced service manual (per unit) should be provided as soft copy and hard copy.

16. Defibrillator protected.